

Creatinine

Jaffe. Colorimetric – kinetic
For in -vitro diagnostic use only.
Store at 2-8°C

PRINCIPLE OF THE METHOD

The assay is based on the reaction of creatinine with sodium picrate as described by Jaffe.
Creatinine reacts with alkaline picrate forming a red complex. The time interval chosen for measurements avoids interferences from other serum constituents.
The intensity of the color formed is proportional to the creatinine concentration in the sample¹.

CLINICAL SIGNIFICANCE

Creatinine is the result of the degradation of the creatine, component of muscles, it can be transformed into ATP, that is a source of high energy for the cells. The creatinine production depends on the modification of the muscular mass, and it varies little and the levels usually are very stable.
Is excreted by the kidneys. With progressive renal insufficiency there is retention in blood of urea, creatinine and uric acid. Elevate creatinine level may be indicative of renal insufficiency^{1,4,5}.
Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

REAGENTS

R 1 Picric Reagent	Picric acid	17.5 mmol/L
R 2 Alkaline Reagent	Sodium hydroxide	0.29 mol/L
CREATININE CAL	Creatinine aqueous primary standard 2 mg/dL	

PRECAUTIONS

R1(Picric acid): Corrosive (C):R35:Causes severe burns.
R2(NaOH): Irritant (Xi): R36/38: Irritating to eyes and skin.S26: In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. S37/39: Wear suitable gloves and eye/face protection. S45: In case of accident or if you feel unwell, seek medical advices immediately.

PREPARATION

Working reagent (WR):
Mix equal volumes of R 1 Picric Reagent and R 2 Alkaline reagent.
The working reagent is stable for 10 days at 15-25°C.

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.
Do not use reagents over the expiration date.

Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 492 nm ≥ 1.80 .

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 492 nm (490-510) - Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

SAMPLES

- Serum or heparinized plasma¹.
Creatinine stability: 24 hours at 2-8°C.
- Urine¹: Dilute sample 1/50 with distilled water. Mix. Multiply results by 50 (dilution factor);
Creatinine stability: 7 days at 2-8°C.

PROCEDURE

- Assay conditions:
Wavelength: 492 nm (490-510)
Cuvette: 1 cm. light path
Temperature 37°C / 15-25°C
- Adjust the instrument to zero with distilled water.
- Pipette into a cuvette:

	Blank	Standard	Sample
WR (mL)	1.0	1.0	1.0
Standard (µL)	--	100	--
Sample (µL)	--	--	100

- Mix and start stopwatch.
- Read the absorbance (A₁) after 30 seconds and after 90 seconds (A₂) of the sample addition.
- Calculate: $\Delta A = A_2 - A_1$.

CALCULATIONS

$$\frac{\Delta A \text{ Sample} - \Delta A \text{ Blank}}{\Delta A \text{ Standard} - \Delta A \text{ Blank}} \times 2 \text{ (Standard conc.)} = \text{mg/dL of creatinine in the sample}$$

Conversion factor: mg/dL x 88.4 = µmol/L.

QUALITY CONTROL

If control values are found outside the defined range, check the instrument, reagents and calibrator for problems.
Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES¹

Serum or plasma:

Male	0,7 - 1,4 mg/dL	= 61.8 – 123.7 µmol/L
Female	0,6 - 1,1 mg/dL	= 53.0 – 97.2 µmol/L
Urine: 15-25 mg/Kg/24 h		
Male	10 - 20 mg/Kg/24 h	= 88– 177 µmol/Kg/24
Female	8 – 18 mg/Kg/24 h	= 71– 177 µmol/Kg/24

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

	Intra-assay(n=20)		Inter-assay(n=20)	
Mean (mg/dl)	1.06	3.58	1.03	3.31
SD	0.22	0.06	0.04	0.06
CV (%)	2.07	1.54	3.97	1.75

Measuring range: From detection limit of 0.09 mg/dL to linearity limit of 15 mg/dL.

If the results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl 9 g/L and multiply the result by 2.

Precision:

Sensitivity: 1 mg/dL = AA 0,03 A/min . mg/dL

Accuracy: Results obtained using ATLAS reagents (y) did not show systematic differences when compared with other commercial reagents (x). The results obtained using 50 samples were the following:

Correlation coefficient (r): 0.986
Regression equation: $y = 0.975x + 0.047$

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

Hemoglobin (1 g/L), Bilirubin (55 mg/dL), interfere¹.
A list of drugs and other interfering substances with creatinine determination has been reported by Young et. al^{2,3}.

NOTES

- Calibration with the aqueous Standard may cause a systematic error in automatic procedures. In these cases, it is recommended to use a serum Calibrator.
- Use clean disposable pipette tips for its dispensation.

BIBLIOGRAPHY

- Murray R.L. Creatinine. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 1261-1266 and 418.
- Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
- Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
- Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
- Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.

ATLAS MEDICAL
Unit 4, William James House
Cowley Rd, Cambridge, CB4 0WX
Tel: ++44 (0) 1223 858 910
Fax: ++44 (0) 1223 858 524

PPI430A01
Revision A (28.02.2007)