



γ -GT


Carboxy substrate kinetic. Liquid

Quantitative determination of gamma-glutamyl transferase (γ -GT)

Store at 2-8 °C

PRINCIPLE OF THE METHOD

Gamma-glutamyl transferase (γ -GT) catalyses the transfer of γ -glutamyl group from γ -glutamyl-p-nitroanilide to acceptor glycylglycine, according to the following reaction:



The rate of 2-nitro-5-aminobenzoic acid formation, measured photometrically, is proportional to the catalytic concentration of γ -GT present in the sample^{1,2}.

CLINICAL SIGNIFICANCE

Gamma-glutamyl transferase (γ -GT) is a cellular enzyme with wide tissue distribution in the body, primarily in the kidney, pancreas, liver and prostate.

Measurements of gamma-glutamyl transferase (γ -GT) activity are used in the diagnosis and treatment of hepatobiliary diseases such biliary obstruction, cirrhosis or liver tumours^{1,2,5,6}.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

REAGENTS

R1 Buffer	TRIS pH 8.6 Glycylglycine	100 mmol/L 100 mmol/L
R2 Substrate	L- γ -glutamyl-3-carboxy-4-nitroanilide	

PREPARATION

Working reagent (WR)

Mix: 4 vol. (R1) Buffer + 1 vol. (R2) Substrate

Stability: 21 days at 2-8 °C

or 5 days at room temperature 15-25 °C

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8 °C, protected from light and contaminations prevented during their use.

Do not use reagents over the expiration date

Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 405 nm \geq 1.20.

ADDITIONA EQUIPMENT

- Spectrophotometer or colorimeter measuring at 405 nm.
- Thermostatic bath at 25°C, 30°C o 37°C (\pm 0.1°C)
- Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

SAMPLES

Serum ¹. γ -GT is stable for at least 3 days at 2-8 °C, 8 hours at 15-25 °C and 1 month at – 20°C.

PROCEDURE

1. Assay conditions:
 Wavelength405nm
 Cuvette.....1 cm light path
 Constant temperature25°C / 30°C/ 37°C.
2. Adjust the instrument to zero with distilled water or air.
3. Pipette into a cuvette (^{note 1})

WR (ml)	1.0
Sample (μ L)	100

4. Mix, wait for 1 minute.
5. Read initial absorbance (A) of the sample, start the stopwatch and read absorbances at 1 minute intervals thereafter for 3 minutes.
6. Calculate the difference between absorbances and the average absorbance differences per minute ($\Delta A/\text{min}$).

CALCULATIONS

$(\Delta A/\text{Min}) \times 1190 = \text{U/L of } \gamma\text{-GT}$

Units: One international unit (IU) is the amount of enzyme that transforms 1 μ mol of substrate per minute. In standsard conditions. The concentration is expressed in units per litre of sample (U/L).

TEMPERATURE CONVERSION FACTORS

To correct results to other temperatures multiply by:

Assay temperature	Conversion factor to		
	25°C	30°C	37°C
25°C	1.00	1.37	1.79
30°C	0.73	1.00	1.30
37°C	0.56	0.77	1.00

QUALITY CONTROL

If control values are found outside the defined range, check the instrument, reagents and technique for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES ¹

	25°C	30°C	37°C
Men	4-18 U/L	5-25 U/L	7-32 U/L
Women	6-28 U/L	8-38 U/L	11-50 U/L

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: Form detection limit of 2U/L to linearity limit of 250 U/L.

If the results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl 9 g/L and multiply the result by 2.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
Mean (U/L)	38.0	188	37.5	190
SD	0.79	2.57	0.96	2.61
CV (%)	2.09	1.36	2.56	1.37

Sensitivity: 1 U/L = 0.0074 ΔA/min.

Accuracy: Results obtained using Atlas reagents (y) did not show systematic differences when compared with other commercial reagents (x).

The results obtained using 100 samples were the following:

Correlation coefficient (r): 0.9960

Regression equation: $y=0.9897x-0.0879$

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

Plasma should not be used ,anticoagulants inhibit the enzyme. Gross haemolysis interferes in the assay¹.

A list of drugs and other interfering substances with γ -GT determination has been reported by Young et. Al 3,4.

BIBLIOGRAPHY.

1. Gendler S. γ-GT. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 1120-1123,
2. Persijn J P et al. J Clin Chem Clin Biochem 1976: (14) 9: 421 – 427.
3. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
4. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
5. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.

6. Tietz N W et al. Clinical Guide to laboratory Tests, 3rd ed AACC 1995.

ATLAS MEDICAL

William James House, Cowley Road, Cambridge

Tel: ++44 (0) 1223 858 910

Fax: ++44 (0) 1223 858 524

PPI464A01

Revision A (14.07.2007)