



LIQUID URIC ACID

For *in-vitro* diagnostic use only.
Store at 2-8°C.

PRINCIPLE OF THE METHOD

Uric acid is oxidized by uricase to allantoin and hydrogen peroxide (2H₂O₂), which under the influence of POD, 4-aminophenazone (4-AP) and 2-4 Dichlorophenol sulfonate (DCPS) forms a red quinoneimine compound:



The intensity of the red color formed is proportional to the uric acid concentration in the sample.

CLINICAL SIGNIFICANCE

Uric acid and its salts are end products of the purine metabolism. With progressive renal insufficiency, there is retention in blood of urea, creatinine and uric acid.

Elevate uric acid level may be indicative of renal insufficiency and is commonly associated with gout^{1,56}

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

REAGENTS

R 1	Phosphate pH 7.4	50 mmol/L
Buffer	2-4 Dichlorophenol sulfonate	4 mmol/L
R 2	Uricase	60 U/L
Enzymes	Peroxidase (POD)	660 U/L
	Ascorbate oxidase	200 U/L
	4-Aminophenazone (4-AP)	1 mmol/L
URIC ACID	Uric acid aqueous primary standard	6 mg/dL

PREPARATION

Working reagent (WR):

Mix equal volumes of R 1 Buffer and R 2 Enzymes.

The working reagent is stable 1 week at refrigerator (2-8°C) or 4 days at room temperature (15-25°C).

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.

Do not use reagents over the expiration date.

URIC ACID CAL Once open is stable up to 1 month when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.

Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 520 nm ≥ 0.16 .

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 520 nm.
- Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

SAMPLES

- Serum or plasma: Stability 3-5 days at 2-8°C or 6 months at -20°C. - Urine (24 h): Stability 4 days at 15-25°C, pH >8. Dilute sample 1/50 in distilled water. Mix. Multiply results by 50 (dilution factor); if urine is cloudy; warm the specimen to 60°C for 10 min to dissolve precipitated urates and uric acid. Do not refrigerate.

PROCEDURE

1. Assay conditions:

Wavelength: 520 nm (490-550)

Cuvette : 1 cm light path

Temperature: 37°C / 15-25°C

2. Adjust the instrument to zero with distilled water.
3. Pipette into a cuvette:

	Blank	Standard	Sample
WR (mL)	1.0	1.0	1.0
Standard (μL)	--	25	--
Sample (μL)	--	--	25

4. Mix and incubate for 5 min at 37°C or 10 min at 15-25°C.

5. Read the absorbance (A) of the samples and Standard, against the Blank. The colour is stable for at least 30 minutes.

CALCULATIONS Serum or plasma

$$\frac{(A)\text{Sample}}{(A)\text{Standard}} \times 6 (\text{Standard conc.}) = \text{mg/dL}$$

uric acid in the sample (A)Standard

Urine 24 h

$$\frac{(A)\text{Sample}}{(A)\text{Standard}} \times 6 \times \text{vol. (dL) urine 24 h}$$

=mg/24 h uric acid.

Conversion factor: mg/dL x 59.5 = μmol/L.

QUALITY CONTROL

Control sera are recommended to monitor the performance of assay procedures. If control values are found outside the defined range, check the instrument, reagents and calibrator for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES

Serum or plasma:

Women 2.5 - 6.8 mg/dL a 149 - 405 μmol/L

Men 3.6 - 7.7 mg/dL = 214 - 458 μmol/L

Urine: 250 - 750 mg/24 h = 1.49 - 4.5 mmol/24 h

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: From *detection limit* of 0.03 mg/dL to *linearity limit* of 25 mg/dL.

If the results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl 9 g/L and multiply the result by 2.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
Mean (mg/L)	4.74	11.4	4.72	11.2
SD	0.03	0.06	0.07	0.15
CV (%)	0.63	0.56	1.58	1.36

Sensitivity: 1 mg/dL = 0.0347 A.

Accuracy: Results obtained using ATLAS reagents (y) did not show systematic differences when compared with other commercial reagents (x). The results obtained using 50 samples were the following:

Correlation coefficient (r): 0.99.

Regression equation: y=1.005x + 0.0005.

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

No interferences were observed to bilirubin up to 170 μmol/L, hemoglobin up to 130 mg/dL and ascorbic acid up to 570 μmol/L².

A list of drugs and other interfering substances with uric acid determination has been reported by Young et. al '4.

NOTES

1. Calibration with the aqueous standard may cause a systematic error in automatic procedures. In these cases, it is recommended to use a serum Calibrator.
2. Use clean disposable pipette tips for its dispensation.

BIBLIOGRAPHY

1. Schultz A. Uric acid. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 1261-1266 and 418.
2. Fossati Pet al. Clin Chem 1980;26:227-231.
3. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
4. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
5. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
6. Tietz N Wet al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.

ATLAS MEDICAL

Unit 4, William James House

Cowley Rd, Cambridge, CB4 0WX

Tel: ++44 (0) 1223 858 910

Fax: ++44 (0) 1223 858 524

PPI203A01

Revision A (04.07.2005)